

## Current status and issues

Things to be aware of at the present time (updated April 2022):

1. The wind blind is currently out of commission, pending repairs to a cable.
2. Information regarding the filters installed in the instrument can be found [below](#).

## Where to find the information you need

This wiki is intended to supplement the existing TCS and instrument user's guides, filling the void of information about the telescope and facility, and providing recipes for observing with the telescope-instrument-TCS combination.

For detailed information about the instruments, please refer to the manuals listed below:

- [SHOC wiki](#)
- "<https://topswiki.sao.ac.za>SHOC User Guide (SHOCnHelpful)"<https://topswiki.sao.ac.za>, by Rocco Coppejans ? contains useful instrument info, but new software described on SHOC wiki
- [SHOC exposure time and prebin calculator](#)
- [SHOC data reduction pipeline](#)
- "<https://topswiki.sao.ac.za>SAAO CCD Camera Control Program, version 1.2"<https://topswiki.sao.ac.za> by J. W. Menzies

Much of this page assumes that the user has never before set foot in the dome, and has no local human to guide them. [Telescope pointing limits](#), [troubleshooting information](#) and [start-up](#) and [shutdown](#) recipes should prove useful for experienced observers and novice users alike.

## Contact numbers

**To receive calls**, the phone in the warm room can be reached by the following numbers:

Extension number: 9108 (to phone from within the observatory)

To phone the dome from outside: 023 100 0222 (+27 23 100 0222 from outside South Africa)

**To make calls** if you have a problem that isn't addressed in this wiki or the [fault forum](#) (remember to dial the \*\* where given):

1.9m/74-inch observer: 9107 (often a good place to start, given that the TCS and SHOC interfaces are common to both telescopes)

Standby electronics technician: \*\*103

Standby mechanical technician: \*\*104

Standby IT technician: \*\*113

Extension number: 9108 (to phone from within the observatory) To phone the dome from outside: 023 100 0222 (+27 23 100 0222 from outside South Africa)

## A guided tour of the dome

The purpose of this section is to introduce you to the facilities and equipment you will need during your run. The tour begins at the front door of the 1-m dome. To aid orientation, the domes of MeerLICHT, Lesedi, and the 40 inch are aligned East-West, perpendicular to the main road that passes SALT at the Northern end of the plateau and leads to the "https://topswiki.saa.ac.za/Robotic Row" of telescopes in the South. A code (which you can obtain from the standby technicians/your support astronomer/other observers) is required to open the door.

### Main telescope power switch

Just inside the entrance on the right-hand wall is the main telescope power key switch. It should always be set to PLC for the telescope to be used. If it is not, call a technician.

### Light switches

- **Foyer and stairwell light switches** can be found on the right-hand wall, just around the corner to the right, at the foot of the stairs. There is a second switch for the stairwell lights at the top of the stairs.
- **Light switches on the left at the top of the stairs.** One is for the stairwell lights; the other for the observing floor incandescents.
- The **switch for observing floor fluorescent lights** is located inside the small door facing the stairs, next to the controls for the air conditioning.
- The **switch for upper dome fluorescent lights** is on the grey Observing Floor Control Panel on the wall by the North pier.

### Ground floor facilities

As you enter the dome on the ground floor, you will find the following facilities:

- **Fire extinguisher:** mounted to the wall on the right as you enter.
- **First aid kit:** on a shelf to the left as you enter the dome.
- **Toilet:** through the first door on the left as you enter the dome. The light switch is outside the door on the right. (Don't be alarmed if you discover that these facilities are occupied during the night; the neighbouring IRSF does not have its own toilet.)
- **Kitchen:** located beyond the staircase on the right. It has recently been refurbished and has a sink with running water, a microwave, coffee maker, kettle, toaster, snackwich and an array of cutlery and cheerful crockery. There are jars of teas, coffees, hot chocolate, sugar and creamer. If these jars become empty, return them to the hostel kitchen in your night lunch bag for refilling. The kitchenware is cleaned daily by the hostel staff, except at weekends.

## The warm room

The warm room was moved downstairs in 2018 - it is the room last room on the left when you enter the dome. Here you have the following resources at your disposal:

### Computing facilities

In the warm room, you will find:

- **For STE3 or STE4**, a monitor at the right-hand end of the desk labelled **imag40**, running the CCD Camera Control software.
- A thin client with monitor labelled **ltsp** that saves the data from the instrument PC and can be used to run a web browser, ds9, etc. **For SHOC**, it can be used to run the instrument control software via a web browser. It is possible to download your data to a laptop on the network from this PC (use **scp** or **sftp ccd40@ltsp.suth** - the monitor is labelled with the password).
- A PC running the **Telescope Control Software (TCS)** for acquisition and guiding. Associated with this is a second monitor which displays the target and telescope coordinates, etc.
- Several network cables and power sockets for use with laptops.
- A black & white printer. Print command: **lpr -Ppr40in filename**

There are no DVD writers, USB ports or DAT drives in the dome for copying data, this should be done using scp.

### Log books, etc.

In the warm room, in labelled boxfiles, you will find the following resources to help with your observations:

- **Users' guides** for the TCS, STE3/STE4, SHOC and the retired UCT CCD, which contains useful information about the online reductions programme for STE3/4.
- A triplicate **CCD Camera log book** in which to record details of each image if you are using STE3/STE4. If you finish a log book, there are more on top of the metal cupboard on the observing floor. Please take the next book in numerical order. There is no such log for use with SHOC; you are advised to keep your own records, e.g. in a spreadsheet.
- A catalogue of photometric **standard stars**.
- An A4 sheet entitled **Sutherland Observing Record**, on which to log the duration of observations, time lost, etc. each night.

### General warm room resources

1. A **stereo** compatible with iPods and iPhones, and plays CDs, MP3s from DVDs and USB devices, and has a radio.
2. Assorted desk **lamps**.
3. A portable **heater** under the desk.

Please do **not** plug lamps or heaters into the UPS sockets!

## The Observing Floor

At the top of the stairs, you will find the telescope on the "https://topswiki.sao.ac.za/observing-floor" https://topswiki.sao.ac.za. The North pier is nearest the stairs. Familiarise yourself with the following:

- **Blue console on the East side of the North pier.**
- **Grey box by the blue console.** If the telescope reaches its safe operating limits, the red light on this box will flash, an alarm will sound and the telescope drives and dome rotation will be disabled. There is an override key attached to the box, used to allow the telescope to be moved out of the limits.
- **Emergency stop button on the grey box.** There is a large red button to press if the drive motors fail to stop. If you have reason to press this, call a technician to investigate.
- **XY-slide reset button on the grey box.** On the top of the grey box with the emergency stop button is a small blue button to press if the TCS XY-slides fail to respond.
- **PC monitor to the right of the North pier.** This screen displays pointing information, including your target coordinates and current telescope coordinates, for use when slewing the telescope manually.
- **Observing Floor Control Panel.** On the North wall by the North pier there is a grey box labelled "https://topswiki.sao.ac.za/Dome Shutter and Windblind Control" https://topswiki.sao.ac.za, which has red buttons for opening/closing the dome shutter and raising/lowering the windblind. One must first switch on the telescope power (blue console) and press the top button labelled "https://topswiki.sao.ac.za/Request Local Control" https://topswiki.sao.ac.za to activate the panel. The red indicator will illuminate when the panel is active. To open/close the dome, push and hold the appropriate button until the shutters stop moving and you hear a beep. To raise/lower the windblind, push and hold the appropriate button until the blind reaches the desired height. **Do not raise the windblind while pointing into the wind; first rotate the dome by 180°, raise the blind then rotate back.** There is also a switch for the upper dome lights.
- **Dome flat lamp switch.** On the North wall, just above and left of the monitor, is a dimmer switch that controls the incandescent lamp used for taking dome flats.
- A red **balance weight** on a rope, which can be hung from an appropriate place below the primary mirror cell if the telescope jumps while tracking (usually when tracking around the zenith).
- Tall and short **wheeled ladders** for hanging the weight and for accessing the reset buttons on the Merlin transputer crates. Make sure the ladders are clear of the telescope while observing (the taller ladder is usually stored on the West side of the observing floor under the catwalk, the other is pushed to the East side of the North pier).

## The Telescope

On the telescope itself, identify the following:

- **Mirror cover motor:** on the East side of the telescope below the primary mirror is a button to open/close the primary mirror covers.
- **Acquisition/guider box and autoguider camera:** the former is a large black box mounted directly below the primary mirror cell. The XY-slides and acquisition/autoguider camera are mounted on the South side of the guider box.
- **TCS Merlin crate:** above the autoguider camera is a small blue electronics crate. This is the Merlin transputer for the TCS. On the side is a red reset button that should be pressed if the TCS encounters a "https://topswiki.sao.ac.za/Merlin error" https://topswiki.sao.ac.za.
- **Filter wheels:** The blue metallic box below the acquisition/guider box contains the filter wheel(s), labelled on the East and West sides.
- **Instrument:** SHOC, STE3 or STE4 (or your own instrument) will be mounted below the filter box.

- **STE3/STE4** has its own blue **instrument Merlin crate** mounted to the cylindrical dewar. There is a red reset button on the side of the crate that should be pressed if the instrument control software hangs, crashes or reports a "https://topswiki.sao.ac.zaMerlin error" https://topswiki.sao.ac.za. There is no need for observers to fill the dewar with liquid nitrogen; this is done by technicians in the morning and afternoon.
- **SHOC** does not require liquid nitrogen; it is peltier-cooled. It has no Merlin crate. The control PC ("https://topswiki.sao.ac.zashocnawe" https://topswiki.sao.ac.za or "https://topswiki.sao.ac.zashocndisbelief" https://topswiki.sao.ac.za) is mounted in a crate below the primary mirror cell.

## Filters

The blue metallic box below the acquisition/guider box contains the filter wheel(s), labelled on the East and West sides. Each wheel has 8 filter positions; position 1 is the reference and position 8 is always empty, so that with two wheels installed, one wheel should always have position 8 in the beam (unless one wheel contains neutral density filters). **Filters must not be removed from the filter wheels.**

Throughout 2023, the filters in the 1-m wheels will be as follows:

Wheel	1	2	3	4	5	6	7	8
Geneva A1	U	B	V	R	I	BG38	Empty	Empty
Geneva B1	u'	g'	r'	i'	z'	H? (20nm)	[O III] 501nm	Empty

## The old warm room

The old warm room is situated on the East side of the observing floor. The light switch is a dial just inside the main door on the right as you enter. The switch for the observing floor fluorescent lights is just inside the hobbit door to this room at the top of the stairs.

## Observing limits

### Meteorological constraints

The dome must be closed when any of the following weather limits is reached:

- **Humidity:**  $T_{\text{external}} - T_{\text{dew}} \geq 1.5$
- **Wind speed:** 60 km/h
- **Heavy cloud**

### Meteorological information

The following should be used for regular monitoring of external conditions while observing:

- DIMM seeing data and all the weather stations on the plateau are summarised on [this page](#)
- For assessing sky conditions, SALT and LCO operate **all sky cameras**, also available on the website above.

## Telescope pointing limits

The operational limits of the telescope are depicted in the figure below.

If the telescope reaches the limits, the drive motors will cut out, dome rotation will be disabled, an alarm will sound and a red light will flash on the grey box on the East side of the North pier. Call a technician to recover the telescope.

. 1.0-m telescope observing limits

## Start-up recipe

Follow this procedure to prepare the telescope and dome for observing:

1. **Check the weather conditions [here](#)**. If windspeed <60km/h and  $T - T_{\text{dew}} > 1.5$  (see [observing limits](#)), continue with step 2.
2. Check the observing floor and **clear any obstacles** (e.g. ladders, liquid nitrogen filler pipe must not be inserted in instrument) that may obstruct the telescope.
3. **Switch off the [lights](#)** everywhere except in the warm room.
4. Check that **Telescope Power Control** (on the wall just inside the front door) is set to **PLC**. If not, call a technician.

If observing in Sutherland: in the warm room, **on the wall-mounted mimic panel** at the left-hand end of the desk (**if observing remotely, these things will be done by the on-site technicians**):

5. Turn **TCS LOCKOUT off** (turn knob anticlockwise).
  6. Check that **SMD is off**.
- . Wall-mounted mimic panel in the warm room, for steps 5-6 if observing from the dome.

### On the TCS:

If a menu is already selected (text appears below the Target, Dome, Focus etc. buttons), click CLEAR to re-activate the menu buttons.

7. **Request TCS control** by clicking the grey "<https://topswiki.sao.ac.za/FLOOR>" button. It will turn blue and say "<https://topswiki.sao.ac.za/TCS>".
8. Switch **Telescope Power on** (red "<https://topswiki.sao.ac.za/Telescope Power is Off>" button turns green).
9. **Open the dome shutters** by clicking the "<https://topswiki.sao.ac.za/DOME>" pull-down menu and selecting "<https://topswiki.sao.ac.za/Shutter OPEN>". The dome indicator on the top right will turn green.

10. **Open the mirror covers** by clicking the "https://topswiki.sao.ac.zaDOME"https://topswiki.sao.ac.za pulldown menu and selecting "https://topswiki.sao.ac.zaMirror OPEN"https://topswiki.sao.ac.za. The mirror indicator on the top right will turn green.
11. Switch **dome to AUTO** by clicking the red "https://topswiki.sao.ac.zaMAN"https://topswiki.sao.ac.za button (top right of display). It will turn green. At times, it is necessary to either rotate the dome or point the telescope to Zenith so that the dome is able to locate its position and avoid the possibility of the dome not correctly following the telescope .
12. Turn **telescope tracking on** by clicking the red "https://topswiki.sao.ac.zaTracking is OFF"https://topswiki.sao.ac.za button. It will turn green.
13. **Initialise autoguider XY-slides** by clicking the "https://topswiki.sao.ac.zaInitialise"https://topswiki.sao.ac.za button on the bottom right.
14. If necessary, **position the windblind** using the appropriate button on the "https://topswiki.sao.ac.zaDOME"https://topswiki.sao.ac.za pulldown menu.

. TCS status panel

## TCS start-up procedure

### If the TCS software is running

If the TCS is up and running (as is usually the case), do the following at the start of the night:

- **Initialise the XY slides** by clicking **Initialize** in the "https://topswiki.sao.ac.zaXY Slides Control"https://topswiki.sao.ac.za panel on the TCS display.
- **Select the appropriate guide mirror position**

The guide mirror must be **out of the beam** for:

- skyflats or domeflats
- while integrating with the instrument PC
- autoguiding

The guide mirror must be **in the beam** while:

- acquiring a field with the acquisition camera

The guide mirror can be moved in and out of the beam by clicking the large button above the "https://topswiki.sao.ac.zaXY Slides Control"https://topswiki.sao.ac.za box on the TCS screen. If the button is red and says "https://topswiki.sao.ac.zaGUIDE MIRROR IS IN BEAM"https://topswiki.sao.ac.za, clicking it will move the mirror, turning the button green and changing the message to "https://topswiki.sao.ac.zaGUIDE MIRROR IS OUT OF BEAM"https://topswiki.sao.ac.za. The converse is also true.

## If the TCS PC is on but the software is not running

If the TCS PC is alive and responding but the programme is not running, it can be started by double-clicking on the TCS icon on the desktop and entering the password in the pop-up box (consult local technical staff for the password).

## If the TCS PC is switched off or not responding

Please call the duty electronics technician.

## If the TCS-web interface is not responding

If the TCS-web is not responding to any inputs, then make sure that the WebUser has Control button is clicked and shows green.

To access the TCS from the web browser, click here TCS-Web.

## Instrument start-up procedure

### STE3/STE4

1. Follow the start-up procedure described in Sections 1 and 2 of the CCD Camera Control Program, version 1.2 <https://topswiki.sao.ac.za>, by J. W. Menzies.

**Important:** Ensure that the following are correct in the blue "https://topswiki.sao.ac.zaDetails of Run" <https://topswiki.sao.ac.za> startup panel:

- **CCD:** select STE3 or STE4
- **Wheel A & B:** enter "https://topswiki.sao.ac.za" <https://topswiki.sao.ac.za> into both boxes to enable filter wheels
- **Next file** number: if this is incorrect, existing images could be overwritten

2. Initialise the filter wheels, move to the desired filter and adjust the prebin or window setting, if required.

3. Take a bias frame (run a single line module with Repeats=1, Image Type=Bias and Save=Yes) and **check that it saves to:**

`/data/40in/imag40/YYYY/MMDD on ltsp.`

(Note: The data is initially written to the local disk on imag40 (`/data/ccd/image/annnxxxx.fits`). An rsync job runs on suthastro every 15 seconds which moves the data onto the SAN. This is accessible from the thin client, logged in as ccd40 in `/data/40in/imag40/YYYY/MMDD`).



## SHOC

Follow the instructions on the [SHOC wiki](#).

## Observing recipes: sky flats

### Sky flats

If sky conditions are photometric during morning/evening twilight, you will be able to take sky flats:

1. Follow [facility start-up procedure](#).
2. Follow the appropriate [instrument setup procedure](#).
3. **Slew the telescope to the zenith**, with the open dome shutter facing in the opposite direction to the rising/setting sun. Leave the drives in FAST mode (i.e. not tracking).
4. On the TCS, move the **guide mirror out of beam**.
5. On the instrument PC, **select the required filter**. If you need flats in a variety of filters, be sure to cycle through them in the correct order (e.g. UBVRi during evening twilight; IRVBU in the morning). This also applies if you need a range of prebin settings (i.e. start with low binning in the evening; high to low binning in the morning).
6. **With STE3/STE4**, click "<https://topswiki.sao.ac.za/Expose>"<https://topswiki.sao.ac.za/>, create a module, select "<https://topswiki.sao.ac.za/Flat>"<https://topswiki.sao.ac.za/> for the image type, and "<https://topswiki.sao.ac.za/Sky>"<https://topswiki.sao.ac.za/> from the pulldown menu beside it. Keep the count rates between 20000 and 50000. Remember to select "<https://topswiki.sao.ac.za/Yes>"<https://topswiki.sao.ac.za/> under "<https://topswiki.sao.ac.za/Save to Disk>"<https://topswiki.sao.ac.za/>. Snapshots can be used to gauge integration times and can be saved to disk if appropriate. Below is a [recipe for taking sky flats](#). See Section 4 of "<https://topswiki.sao.ac.za/SAO CCD Camera Control Program, version 1.2>"<https://topswiki.sao.ac.za/> by J. W. Menzies for more detailed instructions on setting up a module.
7. Take a **minimum of 15 flats in each filter/prebin setting**, keeping exposure times  $\geq 1$ s with STE3/4 to allow the shutter time to open and close (SHOC has no shutter so does not have this limitation).

### Sky flat recipe for STE3/4: Modules

Since all exposures with STE3/4 must be performed via a module (except for snapshots), it may seem challenging to set up and edit modules quickly enough to obtain sufficient sky flats during twilight. However, you can do this pretty quickly using the following example:

1. **Before twilight**, set up a module for your first filter and an exposure time of 1s (do not go below 1s), with 1 repeat.
2. For evening twilight, perform [steps 1-11 of the start-up recipe](#) to setup the telescope and facility prior to sunset.
3. Click the small grey "<https://topswiki.sao.ac.za/On>"<https://topswiki.sao.ac.za/> button below the CCD image display.
4. Shortly before (morning) or after (evening) sunrise/set ~10 minutes after sunset for B) click "<https://topswiki.sao.ac.za/Snapshot>"<https://topswiki.sao.ac.za/> ? Exposure time = 1s. Click "<https://topswiki.sao.ac.za/Ready>"<https://topswiki.sao.ac.za/> to expose.
5. Hover the cursor over the top portion of the image display as the image reads out, and read off the count level in the yellow "<https://topswiki.sao.ac.za/ADU>"<https://topswiki.sao.ac.za/> box below the image.

Assess how the exposure time needs adjusting to maintain counts of 20000-50000 as the sky darkens/brightens.

6. **During readout** click "https://topswiki.sao.ac.zaExpose" https://topswiki.sao.ac.za ? acknowledge the error message ? "https://topswiki.sao.ac.zaSelect" https://topswiki.sao.ac.za module. Click on the desired module ? "https://topswiki.sao.ac.zaEdit" https://topswiki.sao.ac.za. Adjust the exposure time in the box at the bottom ? "https://topswiki.sao.ac.zaReplace" https://topswiki.sao.ac.za ? "https://topswiki.sao.ac.zaSave" https://topswiki.sao.ac.za ? "https://topswiki.sao.ac.zaContinue" https://topswiki.sao.ac.za.
7. Enter number of modules = 1, Image Type = Flat ? Sky. Save to disk = Yes.
8. When the CCD finishes reading out, click "https://topswiki.sao.ac.zaReady" https://topswiki.sao.ac.za.
9. Repeat steps 5-8 until you have enough flats, or until the counts are too low in that filter.
10. To move onto the next filter, repeat step 6, but edit the filter number as well as the exposure time.
11. Continue steps 5-9 until you have sufficient flats in all your program filters.

With practice, you will easily be able to do this before the CCD finishes reading out in 2x2 binning.

## Observing recipes: on sky

The pointing model of the 1.0-metre is greatly improved in recent years, but is not quite perfect. It may be necessary to point to a bright star near your target to allow the TCS to calculate the pointing offset in each part of the sky. One set of offsets will be applicable to a range of positions.

### Determining pointing offsets using a bright star

If you have trouble finding your target -- and you have checked the equinox and coordinates have been entered correctly into the TCS -- the problem may be the pointing. An offset can be determined for that part of the sky using the following procedure:

1. Follow the facility start-up procedure.
2. Using the Astronomical Almanac or Sutherland Almanac, select a bright star (~4th mag) near your science target (ideally within RA~30m and dec~10°).
3. Check that the target and bright star selected are within the observing limits.
4. **Initialise** the XY-slides.
5. Move the **guide mirror into the beam**.
6. Type the **equinox** of the bright star coordinates into the "https://topswiki.sao.ac.zaSET EQUINOX" https://topswiki.sao.ac.za box and press "https://topswiki.sao.ac.zaENTER" https://topswiki.sao.ac.za on the keyboard.
7. Click "https://topswiki.sao.ac.zaTARGET" https://topswiki.sao.ac.za and type the RA & Dec of the bright star into the appropriate boxes, then click "https://topswiki.sao.ac.zaAuto" https://topswiki.sao.ac.za for automatic slewing. If nothing happens, check that you have completed the start-up procedure. The auto slew can be cancelled by clicking the "https://topswiki.sao.ac.zaManual" https://topswiki.sao.ac.za or "https://topswiki.sao.ac.zaStop Tel. + Dome" https://topswiki.sao.ac.za button.
8. Set "https://topswiki.sao.ac.zaExp Time" https://topswiki.sao.ac.za to ~1s and "https://topswiki.sao.ac.zaWINDOW" https://topswiki.sao.ac.za to "https://topswiki.sao.ac.zaFull frame" https://topswiki.sao.ac.za to take images with the acquisition camera.
9. Use the arrow buttons in the lower left of the TCS display for fine adjustment of the telescope, while watching the image display to centre the bright star. If the star doesn't appear, first check the equinox and coordinates are correct, then adjust "https://topswiki.sao.ac.zaExp Time" https://topswiki.sao.ac.za to

- ~10s and look for light from the star beaming in from one side of the image, revealing its position.
10. Once the bright star appears, select the minimum exposure time, centre the star on the image and click "https://topswiki.sao.ac.zaP" https://topswiki.sao.ac.za by the Target button.
  11. Select "https://topswiki.sao.ac.zaZERO POINTS" https://topswiki.sao.ac.za ? "https://topswiki.sao.ac.zaEnter RA & Dec of Star" https://topswiki.sao.ac.za, then type the **catalogue coordinates** of the bright star into the boxes and press "https://topswiki.sao.ac.zaSET" https://topswiki.sao.ac.za.
  12. The RA & Dec zeropoints will be displayed in the message box. Note them down with the HA & Dec as you will need to re-enter them using "https://topswiki.sao.ac.zaZERO POINTS" https://topswiki.sao.ac.za ? "https://topswiki.sao.ac.zaEnter Zeros Directly" https://topswiki.sao.ac.za if the TCS software is restarted, or if you return to this part of the sky after using different zeropoints in a different position. You do not need different zeropoints for different equinoxes. There is no need to note the collimation coefficient as it is ignored by the pointing model.

## Acquiring a target

1. Follow the facility start-up procedure.
2. Check that the selected target is within the observing limits.

### On the TCS:

1. **Initialise** the XY-slides.
2. Move the **guide mirror into the beam**.
3. Type the **equinox** of the target coordinates into the "https://topswiki.sao.ac.zaSET EQUINOX" https://topswiki.sao.ac.za box and press "https://topswiki.sao.ac.zaENTER" https://topswiki.sao.ac.za on the keyboard.
4. Click "https://topswiki.sao.ac.zaTARGET" https://topswiki.sao.ac.za and type the RA & Dec of the target into the appropriate boxes, then click "https://topswiki.sao.ac.zaAuto" https://topswiki.sao.ac.za to slew. (If you cannot see a "https://topswiki.sao.ac.zaTARGET" https://topswiki.sao.ac.za button, click the "https://topswiki.sao.ac.zaO" https://topswiki.sao.ac.za in the left-hand corner below the "https://topswiki.sao.ac.zaSET EQUINOX" https://topswiki.sao.ac.za box.)
5. Set "https://topswiki.sao.ac.zaExp Time" https://topswiki.sao.ac.za to ~2-4s and "https://topswiki.sao.ac.zaWINDOW" https://topswiki.sao.ac.za to "https://topswiki.sao.ac.zaFull frame" https://topswiki.sao.ac.za to take images with the acquisition camera.
6. Use the arrow buttons in the bottom left of the display for fine adjustment of the telescope while watching the image display to centre the target. Comparison with an 8x8 arcmin finderchart is most helpful.
7. The position should be accurate if you have followed this procedure, so check your zeropoints. If not, adjust "https://topswiki.sao.ac.zaExp Time" https://topswiki.sao.ac.za to search for patterns in fainter stars to compare with your finderchart. If this fails, then reduce the exposure time and scan the field by systematically adjusting RA & Dec.
8. Once the target is identified, move the guide mirror out of beam and position the target on the centre of the science CCD (STE3/4 or SHOC).
  - **For STE4**, position the star approximately on coordinates (210, 270) of the TCS image display, to coincide with the centre of the science detector (for **STE3**, approximately (180,190)).
  - Move the **guide mirror out** to direct the beam to the science detector.

**On the instrument PC:**

- Take a snapshot (STE3/4) or preview (SHOC) to check the positioning of the target on the detector and adjust using the arrow buttons on the TCS.
- Using **STE3/4**, video mode can be invoked by clicking ["https://topswiki.sao.ac.zaSnapshot"](https://topswiki.sao.ac.zaSnapshot)<https://topswiki.sao.ac.za>, entering an integration time, selecting ["https://topswiki.sao.ac.zaVideo"](https://topswiki.sao.ac.zaVideo)<https://topswiki.sao.ac.za> and clicking ["https://topswiki.sao.ac.zaReady"](https://topswiki.sao.ac.zaReady)<https://topswiki.sao.ac.za>. It can be stopped by clicking ["https://topswiki.sao.ac.zaSnapshot"](https://topswiki.sao.ac.zaSnapshot)<https://topswiki.sao.ac.za> and ["https://topswiki.sao.ac.zaStop"](https://topswiki.sao.ac.zaStop)<https://topswiki.sao.ac.za>. Video mode automatically prebins the CCD, so remember to change back to 1x1 afterwards, if appropriate.
- **For a short run of short exposures**, you may set up a module (STE3/4 - see ["https://topswiki.sao.ac.zaSAAO CCD Camera Control Program, version 1.2"](https://topswiki.sao.ac.zaSAAO%20CCD%20Camera%20Control%20Program,%20version%201.2)<https://topswiki.sao.ac.za> by J. W. Menzies) and begin to observe

or

- **For a long run or for exposures >100s, you should locate a guide star and set up the autoguider before integrating on your target.**

**Finding a guide star & autoguiding**

On the TCS:

1. Move the **guide mirror out of beam**.
2. In ["https://topswiki.sao.ac.zaExposures & Guiding"](https://topswiki.sao.ac.zaExposures%20&%20Guiding)<https://topswiki.sao.ac.za> set ["https://topswiki.sao.ac.zaExpTime"](https://topswiki.sao.ac.zaExpTime)<https://topswiki.sao.ac.za> ~2s (adjust as necessary).
3. Ensure that the XY slides have been initialised (click ["https://topswiki.sao.ac.zaInitialise"](https://topswiki.sao.ac.zaInitialise)<https://topswiki.sao.ac.za> under XY slides control).
4. Open the guide star software (the icon of a star with a long tail).
5. Enter coordinates of the field and click ["https://topswiki.sao.ac.zaGET IMAGE"](https://topswiki.sao.ac.zaGETIMAGE)<https://topswiki.sao.ac.za>. A window should appear displaying a DSS image.
6. Click on a bright star outside of the green circle in the DSS image. NOTE: this chart is centred on the given coordinates. If you have adjusted the position of the target to be away from the centre of the acquisition frame, there will be the same offset in the selected guide star.
7. Under ["https://topswiki.sao.ac.zaXY Slides Control,"](https://topswiki.sao.ac.zaXY%20Slides%20Control)<https://topswiki.sao.ac.za> click ["https://topswiki.sao.ac.zaGoTo GS"](https://topswiki.sao.ac.zaGoToGS)<https://topswiki.sao.ac.za>. The star should appear in the acquisition image. If not, try a different star.
8. Under ["?Pointer"](https://topswiki.sao.ac.zaPointer)<https://topswiki.sao.ac.za> click ["https://topswiki.sao.ac.zaGuide Star"](https://topswiki.sao.ac.zaGuideStar)<https://topswiki.sao.ac.za> , then click on the chosen guide star in the acquisition camera FOV. This will put a blue box around the star and begin guiding. NOTE: if the star is near a hot pixel, that could cause trouble with guiding. Click ["https://topswiki.sao.ac.zaMove XY"](https://topswiki.sao.ac.zaMoveXY)<https://topswiki.sao.ac.za> and click on the star: this will center it (just moving the XY stages, and not affecting the science target).
9. To stop or restart guiding (e.g. after brief intervals of cloud), press GUIDE.
10. To change exposure times, select exposure time from the drop down menu under Expose.

## TCS operations at the end of an observation

### At the end of an observation:

- Click "https://topswiki.sao.ac.zaInitialize" in "https://topswiki.sao.ac.zaXY Slides Control".
- In "https://topswiki.sao.ac.zaExposures & Guiding" set "https://topswiki.sao.ac.zaWINDOW" to "https://topswiki.sao.ac.zaFull Frame".
- Move "https://topswiki.sao.ac.zaGUIDE MIRROR" into beam to acquire next target.

### At the end of the night:

- In "https://topswiki.sao.ac.zaXY Slides Control", click "https://topswiki.sao.ac.zaSTOP" to cease A/G camera integrations.
- In "https://topswiki.sao.ac.zaXY Slides Control", click "https://topswiki.sao.ac.zaInitialize" to home the XY slides.

## Shutdown procedures

### Closing in bad weather

If any of the weather limits is reached during the night, **perform these steps first, in this order:**

1. **Close mirror covers** ("https://topswiki.sao.ac.zaDOME" pulldown menu, "https://topswiki.sao.ac.zaMirror CLOSE"). Check that mirror indicator turns red & shows "https://topswiki.sao.ac.zaCLOSED".
2. **Close dome shutters** ("https://topswiki.sao.ac.zaDOME" pulldown menu, "https://topswiki.sao.ac.zaDome CLOSE"). Check that dome indicator turns red & shows "https://topswiki.sao.ac.zaCLOSED".
3. Continue with the remaining steps below.

### Normal shutdown

#### On TCS PC:

1. **Stop autoguider exposures** ("https://topswiki.sao.ac.zaSTOP" on bottom left of display)
2. **Initialise XY-slides** ("https://topswiki.sao.ac.zaINITIALISE" on bottom left)
3. **Turn off tracking** ("https://topswiki.sao.ac.zaTracking is OFF" button turns red, green "https://topswiki.sao.ac.zaSLOW" indicators turn red to "https://topswiki.sao.ac.zaFAST")
4. **Park telescope** (click "https://topswiki.sao.ac.zaTARGET" then "https://topswiki.sao.ac.zaPARK"). Ensure telescope reaches HA = -1h, Dec = 0°

5. **Close mirror covers** ("https://topswiki.sao.ac.zaDOME"https://topswiki.sao.ac.za pulldown menu, "https://topswiki.sao.ac.zaMirror CLOSE"https://topswiki.sao.ac.za). Check that mirror indicator turns red & shows "https://topswiki.sao.ac.zaCLOSED"https://topswiki.sao.ac.za
6. **Close dome shutters** ("https://topswiki.sao.ac.zaDOME"https://topswiki.sao.ac.za pulldown menu, "https://topswiki.sao.ac.zaDome CLOSE"https://topswiki.sao.ac.za). Check that dome indicator turns red & shows "https://topswiki.sao.ac.zaCLOSED"https://topswiki.sao.ac.za
7. **Park dome** ("https://topswiki.sao.ac.zaDOME"https://topswiki.sao.ac.za pulldown menu, "https://topswiki.sao.ac.zaDome PARK"https://topswiki.sao.ac.za)
8. **Switch dome to manual** ("https://topswiki.sao.ac.zaAUTO"https://topswiki.sao.ac.za button on top right turns red and shows "https://topswiki.sao.ac.zaMAN"https://topswiki.sao.ac.za)
9. Lower wind blind (if appropriate, from "https://topswiki.sao.ac.zaDOME"https://topswiki.sao.ac.za pulldown menu)
10. Switch **Telescope Power OFF** using red power button

If observing in Sutherland (if working remotely this will be done by on-site technicians), **on wall-mounted mimic panel:**

11. Turn **Lockout ON** (turn switch clockwise)
12. Check that **SMD** is **off**

#### **On instrument PC:**

13. Stop exposures
14. SHOC: turn off camera
15. STE3/4: leave the software running

## **STE3/STE4: CCD Camera log book**

If using STE3 or STE4, please fill in the triplicate CCD Camera log book for each image obtained, pressing hard enough for your words to register on all three sheets. If you are using SHOC, there is no such log so you should note all the relevant information in e.g. a spreadsheet for your records.

The fields are explained as follows:

- Telescope: 1.0-m.
- Focus: This is telescope focus. You may wish to make a note of the acquisition camera focus for your own purposes.

- **CCD Type:** This will be STE3 if the unbinned detector dimensions are 512x512 pixels, else STE4 for 1024x1024.
- **CCD temp:** This refers to the "https://topswiki.sao.ac.zaCu Block" https://topswiki.sao.ac.za temperature reported in the "https://topswiki.sao.ac.zaTemperature" https://topswiki.sao.ac.za section of the instrument status panel of the STE3/4 CCD Camera Control software.
- **Filters:** Enter the filter in each position of the two wheels.
- **Date and Observers:** are self-explanatory.
- **Run number:** This is a 3-digit number that will apply to your data for the duration of your observing run. It will be prefixed to your image files, which will be named annnxxxx.fits, where nnn is the run number, xxxx is the file number. To determine your run number, check the previous run number in the log book and add 1.
- **File No.** is the number of the image (xxxx where the filename is annnxxxx).
- **Object:** target name.
- **RA & Dec:** it may help you to write here the exact telescope coordinates when pointing to your target, as opposed to the catalogue coordinates. This information, together with the associated zeropoints, will enable you to re-acquire the target quickly in future.
- **Equinox:** of your target coordinates.
- **Start UT:** is reported in the status panel of the CCD Camera Control software, and remains visible until the next exposure is started.
- **Start sidereal time:** can be read from the TCS coordinate display monitors.
- **Int Time:** the exposure time in seconds.
- **Preflash:** this is not required with the current detectors. This box can be used to record the prebin setting.
- **Filter and Seeing:** are self-explanatory.
- **Guide star X Y:** these are the XY-slide coordinates of the guide star used for each observation. Read them from the TCS "https://topswiki.sao.ac.zaXY Slides Control" https://topswiki.sao.ac.za window while guiding, as they are not necessarily identical to the values you read from DS9.
- **Remarks:** Any applicable notes, e.g. the RA & Dec zeropoints for the specific coordinates used.

## Troubleshooting

The table below describes some problems you might encounter. Some can be tackled by the observer (see column 4). If the suggested steps to be followed by an observer fail, please call a technician.

Symptom	Cause	Solution	Who to call
The TCS GUI disappears/won't open	The 40inch_lodestar_source_extract service needs to be running on ioserver1.	Check if the process is running, if it is reported as "https://topswiki.saa.ac.za.inactive (dead)"https://topswiki.saa.ac.za then restart it and check again:  ssh io@ioserver1.suth systemctl status 40inch_lodestar_source_extract systemctl start 40inch_lodestar_source_extract systemctl status 40inch_lodestar_source_extract	IT or electronics standby.
Telescope pointing problems	There may be a time server error.	Check the time and date displayed on the TCS. If it is incorrect, phone for assistance.	Electronics standby.
Dome Shutter obscuring the telescope	The dome does not find its position, and therefore does not follow the telescope correctly.	Point the telescope to zenith (with dome in auto), or rotate the dome so that it can establish its position by passing through some of the cardinal points.	Observer

## Webcam

There is a webcam display that should be visible at all times for viewing and listening to the telescope while observing.

Login to the Foscam app with the following details:

**Email:** it@saa.ac.za

**PW:** Ask IT

Download and install Foscam app from here:

[App Store](#)

[Android](#)

**URL:** <https://40incam.suth.saa.ac.za> <-- This only works from Internet Explorer

**Username:** operator

**PW:** 40incam